

URIT

URIT-12 Hemoglobin Meter



- ▶ **Fast hemoglobin and HCT test within 10 seconds**
- ▶ **Accurate & Reliable Result**
- ▶ **Easy 3- step operation**

Global Diagnostics Supplier

Specifications

Parameter: Hemoglobin, HCT (Hematocrit)

Principle: Optical reflectance

Sample: Capillary or venous whole blood (13~15 ul)

Strip: H12 Hemoglobin Test Strip

Speed: ≤ 10 seconds

Measuring Range: 4.0g/dL~24.0g/dL(g/L optional)

Memory: 250 test results

Quality Control: HQ-A Hemoglonbin control solution

Calibration: Factory calibrated according to ICSH

Automatization: Self-checking, auto-judging and displaying malfunction



Power: DC 6V(Two CR2032 Batteries)

Dimension: 102mm×50mm×19mm



Accuracy: accuracy ≥ 96%, the correlation is very good in the comparison between URIT-12 and ICSH method.

Precision: CV < 5%, testing samples including high, medium & low concentration, the CV of 10 repeated test results is 1.8%, 2.2%, 2.7% respectively.



Operation



STEP 1

Turn on the meter, insert the test strip



STEP 2

Apply blood sample to the test strip



STEP 3

Read the test result

URIT MEDICAL ELECTRONIC (GROUP) CO., LTD.

Add: No.4 EAST ALLEY, JIUHUA ROAD, GUILIN, GUANGXI 541001, P.R.CHINA
 Tel: +86 773 2288583/2288589 Fax: +86 773 2288560/2288559
 E-mail: export@uritest.com service@uritest.com Http://www.urit.com

