

Computer requirements for the Aspel products

Aspel S.A.

Computer compatible with IBM (acc. to PN-EN 60950)

Aspel guarantees proper software's operation only on the computers with specification mentioned below.

Product	Operating system	Number of ports		Additional computer configuration	Monitor's resolutions	Soft version
		RS-232	USB			
HolCARD 24W Alfa System A702	Windows XP HE SP2 32/64bit Windows XP Prof SP2 32/64bit Windows 7 32/64bit Windows 8 32/64bit	--	4	CF card reader, CD/DVD R/W	1280x800 1280x1024 1440x900 16000x900	8.00.00
HolCARD 24W Alfa System A812	Windows XP HE SP2 32/64bit Windows XP Prof SP2 32/64bit Windows 7 32/64bit Windows 8 32/64bit	--	4	CF card reader, CD/DVD R/W	1280x800 1280x1024 1440x900 16000x900	8.00.00
HolCARD CR-07 Alfa System	Windows XP HE SP2 32/64bit Windows XP Prof SP2 32/64bit Windows 7 32/64bit Windows 8 32/64bit	---	3	CD/DVD R/W	17" 1024x768 (*)	3.00.00
CardioTEST Alfa System B612	Windows XP HE SP2 32/64bit Windows XP Prof SP2 32/64bit Windows 7 - 32/64bit Windows 8 32/64bit	1	3	CD/DVD R/W, internal or external speaker	17" 1024x768 (*)	3.00.00
CardioTEST Alfa System B612	Windows XP HE SP2 32/64bit Windows XP Prof SP2 32/64bit Windows 7 - 32/64bit Windows 8 32/64bit	1	3	CD/DVD R/W, internal or external speaker	17" 1024x768 (*)	3.00.00
CardioTEST Alfa System A500	Windows XP HE SP2 32/64bit Windows XP Prof SP2 32/64bit Windows 7 32/64bit Windows 8 32/64bit	1	3	CD/DVD R/W, internal or external speaker	17" 1024x768 (*)	3.00.00
CardioTEKA - software v.201	Windows XP HE SP2 32/64bit Windows XP Prof SP2 32/64bit Windows 7 32/64bit	0 or 1 when RS socket is in device	2	CD/DVD R/W, Ethernet card	17" 1024x768 (*)	4.11.00
CardioTEKA AC (AsCARD Coral, AsCARD Amber)	Windows XP HE SP2 32bit Windows XP Prof SP2 32bit Windows 7 32/64bit Windows 8 32/64bit	0	2	CD/DVD R/W, Ethernet and sound cards	17" 1024x768 (*)	4.12.00
CardioTEKA Viewer	Windows XP HE SP2 32bit Windows XP Prof SP2 32bit Windows 7 32bit	0 or 1 when RS socket is in device	2	---	17" 1024x768 (*)	3.24.00

(*) – minimum size and resolution of the monitor

Aspel S.A. is not responsible for not working and improper working of the Aspel's software as well as for any consequences from not working and improper working resulting from noncompliance with the operation manual, using unlicensed software or any other unsure software installed in user's PC where Aspel's software is installed